

Certificate of Compliance



No. 0D160414.TSUU34

Technical Construction File no. TP-16000431-1, TP-16000431-2

Certificate's
Holder:

TOPAIR SYSTEMS, INC.
3182 Monterey Dr. Merrick, NY 11566 USA

Certification ECM
Mark:



Product:
Model(s):

PCR
PCR-060-UV, PCR-090-UV, PCR-120-UV,
PCR-060-HEPA, PCR-090-HEPA,
PCR-120-HEPA

Verification to:

Standard:
EN 61010-1:2010, EN 61326-1:2013

related to CE Directive(s):
2014/35/EU (Low Voltage)
2014/30/EU (Electromagnetic Compatibility)

Remark: The product(s) has been verified on a voluntary basis. The product(s) satisfies the requirements of the Certification Mark of ECM, in reference to the above listed Standard(s). The above Compliance Mark can be affixed on the product(s) accordingly to the ECM regulation about its release and its use. The regulation can be found at www.entecerma.it. This Certificate of Compliance can be checked for validity at www.entecerma.it

This verification doesn't imply assessment of the production of the product(s).

Additional information, clarification about the **CE** Marking:



We attest that a TCF for the **CE** Marking process is in place. Whereas the Manufacturer is Responsible to start the **CE Marking Certification Procedure** and to perform all the necessary activities, as required by the Directive before placing the **CE** Mark on the product(s).

Date of issue 14 April 2016

Expiry date 13 April 2021

Chief Manager
Tim Mahan



Deputy Manager
Viola Miller

