

CERTIFICATE OF TYPE TESTING IN ACCORDANCE WITH EN 14175, PART 3

CERTIFICATE & REPORT NO: INV/EN14175/1010

DATE: 7th June 2023

Fume Hood Manufacturer:
Topair Systems INC
300 First Avenue, Suite 102
Needham
MA 02494
USA



Fume Hood Model:
FH-200
Fume Hood Type:
200cm wide bench-type

External Dimensions:
Height = 2310mm
Width = 2000mm
Depth = 835mm
Internal Dimensions:
Height = 1170mm
Width = 1800mm
Depth = 640mm (wall-sash)
Depth = 585mm (baffle-sash)

Test Opening:
Width: 1800mm
Height: 500mm

Fume Hood Flow:
Volume flow rate: 1750m³/hr (+/-3%)
Face velocity: 0.50m/s (+/-5%)

Fume Hood Containment:
Inner-plane containment:
C1: <0.010ppm at all locations
Outer-plane containment:
C2, C3, C4, C5: <0.010ppm
Robustness of containment:
CR: 0.016ppm (+/-30%)

This is to certify that the fume hood described above has been type tested in accordance with Part 3 of EN 14175, in compliance with the requirements of Part 2 and with reference to Part 1, and resulted in performance characteristics given in test report, INV/EN14175/1010.

Tested and Certified by:
Invent UK Ltd, 85 Southdown Road,
Harpenden, Herts AL5 1PR
t: +44 (0) 7785 245077
email: bicen@invent-uk.com

Dr A F Bicen

A handwritten signature in blue ink that reads 'A.F. Bicen'.

